




Certificate of Traceability

Level I Laboratory Certification

This certifies that **LKF - Laboratorium fuer Klinische Forschung**, using **Roche cobas c501** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#13**

The system evaluated was:

Instrument: Cobas c501	Calibrator Lot: 492058	Application: (V-HbA1c/V-HB)*91.5 + 2.15 Whole Blood
Reagent Lot: 625162, 592643	Calibrator Assigned Values: 0.137, 0.352, 0.697, 0.972, 1.328, 1.897 mmol/L HbA1c; 1.413, 7.621 mmol/L Hb	

Date of Certification: October 1, 2022

Certification Expires: October 1, 2023

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor